

# Simultaneous estimation of spatial and kinetic parameters in single rotation dynamic myocardial SPECT

Jonathan S. Maltz<sup>†‡</sup>, Elijah Polak<sup>‡</sup> and  
Thomas F. Budinger<sup>†‡</sup>

<sup>†</sup>Center for Functional Imaging, LBNL

<sup>‡</sup>Department of Electrical Engineering and Computer  
Sciences, University of California at Berkeley



# Outline

- Problem statement
- Previous work in simultaneous spatio-kinetic estimation in dynamic ECT
- Our approach
- Results
- Conclusions


## Dynamic SPECT problem: Reconstruction directly from projections

- Source distribution changes as camera rotates
- Incomplete spatio-temporal sampling of Radon transform
- Reconstruction requires estimation of both:
  - Geometry of the source distribution
  - Parameters of tracer kinetic model
- Very difficult, ill-conditioned, non-linear problem
- Cannot be assured of a unique solution owing to fundamental nature of problem

# Imaging geometry



# Interpolation of sinotimogram




## Previous work

- Many algorithms have been developed over the past two decades
  - None has satisfactorily addressed the problem when:
    - The projection data contain realistic levels of noise
- AND
- Several regions of differing kinetics are present among the sources


## Simultaneous spatial and kinetic parameter estimation:

- We extend the joint spatial and kinetic parameter estimation approach of Chiao et al. (1994) to multiple dynamic regions




Chiao, Rogers, Clinthorne, Fessler and Hero, IEEE TMI 13(2) 1994


Extension to case of multiple regions is non-trivial:




Normal myocardium



Defect








$$p(\theta(t), s) = i(t) * \underbrace{\sum_{m=1}^3 w_m(\theta, s, k_1^m) e^{-k_2^m t}}_{\text{exponential sum}}$$

# Parameter redundancy in exponential sums



$$f_1 = 6e^{-t/2} + 12e^{-t/8}$$

$$f_2 = 11e^{-t/3} + 7e^{-t/10}$$

# Chief objectives in present approach:


## 1. Ensure robustness against noise:

- Employ highly constrained spatial model - cannot resolve conventional pixel-based image at low SNR's
- Employ global-local hybrid optimization algorithm

## 2. Improve condition of problem:



- Exponential sum problem: Construct new basis set for TAC's to reduce parameter redundancy
- Constrain both geometric and kinetic parameters to feasible ranges
- Reduce number of parameters

## Non-overlapping ellipse ring model





## Orthogonal basis functions

1. Choose range of physiologically feasible rate constants:  
Here we choose  $k_2 \in [0.001, 1] \text{ min}^{-1}$
2. Use singular value decomposition (SVD)  
to find an orthogonal basis for these exponentials
3. Choose first few principal components as approximating  
basis function set
4. Convolve these with the blood input function



## Orthogonal basis functions

1. Choose range of physiologically feasible rate constants:  
Here we choose  $k_2 \in [0.001, 1] \text{ min}^{-1}$
2. Use singular value decomposition (SVD)  
to find an orthogonal basis for these exponentials
3. Choose first few principal components as approximating  
basis
4. Convolve these with the blood input function





Original TAC composition for region  $n$ :

$$f_n(t) = \sum_{m=1}^M k_1^{mn} i(t) * \underbrace{e^{-k_2^{mn} t}}_{\text{original exponential basis}}$$

Approximating discrete TAC:

$$\hat{f}_n[l] = \sum_{m=1}^{\tilde{M}} \mu^{mn} i(t) * \underbrace{u^m[l]}_{\text{orthogonal basis set}} \quad l = 0, 1, \dots, L - 1.$$

Original TAC composition for region  $n$ :

$$f_n(t) = \sum_{m=1}^M k_1^{mn} i(t) * \underbrace{e^{-k_2^{mn} t}}_{\text{original exponential basis}}$$

Approximating discrete TAC:

$$\hat{f}_n[l] = \sum_{m=1}^{\tilde{M}} \mu^{mn} i(t) * \underbrace{u^m[l]}_{\text{orthogonal basis set}} \quad l = 0, 1, \dots, L - 1.$$

Under new basis we enjoy:

- Reduction in parameter redundancy  
 $\Rightarrow$  Better conditioned problem
- Dimensionality reduction ( $\tilde{M} \ll M$ )  
 $\Rightarrow$  Skirts ‘curse of dimensionality’  
 $\Rightarrow$  Lowers computational cost

BUT

TAC’s may now be negative - need to constrain the  $\mu^{mn}$

## Estimation problem:

Simultaneously determine:

- Parameters of ellipse ring model
- Parameters of the compartmental model of each region

Given:

- The measured sinogram
- A rough initial placement of the ring within image space


# Hybrid optimization algorithm

Least squares cost function has thousands of suboptimal local minima

⇒ Multistart algorithm is **essential**

**Starting point generation:** Adaptive simulated annealing (ASA) algorithm


**Parameter value refinement:** Projected descent algorithm - interval bounds on parameters




●	Stochastically generated starting points
→	Local optimisation algorithm trajectories
○	Equal cost contours of objective function
- - - -	Boundaries of basins of attraction



## Algorithm evaluation


- Generated 250,000 photon counts per slice
- Ignore attenuation, scatter & imperfect system response
- Fit to sinogram of segmented annulus phantom -  
Can ellipse model cope with spatial model mismatch?
- Include exaggerated number of compartments in phantom




































Progress (in image space) at generated state: 4112000



Total counts: 2.5e+05



# Results

- This particular configuration tested over 20 noise realizations
- Good visual fits are obtained
- Adjacent regions with similar TAC's may be merged

Metric	Value $\pm$ std. dev.
Mean RMS deviation in recovered TAC's:	6.6 $\pm$ .9%
Mean RMS error in wash-in parameter ( $k_1$ ):	20.8 $\pm$ 4.7%
Mean RMS error in wash-out parameter ( $k_2$ ):	29.6 $\pm$ 4.7%

## Conclusions

- Performance is encouraging - estimates of useful accuracy are obtained
- Execution is slow (typically 12 hours on Pentium II 400MHz)  $\Rightarrow$  need to parallelize algorithm
- Noise and model mismatch problems are minor compared to problem of extricating individual TAC's from their sums
- Algorithm is believed to be the first to successfully perform simultaneous spatial and kinetic estimation on multiple regions at realistic noise levels
- How will non-ideal imaging conditions affect performance?